



Office of the Addl. DHA(M&TB)
South Delhi Municipal Corporation
E-block, 18th floor, Dr. SPM Civic Center,
JLN Marg, New Delhi 110002

No. Addl. DHA/M&TB//SDMC/2018/. 376

Date: 17/10/2018


Amendment

In continuation of our NIT bearing no. Addl. DHA/M&TB/11 dated 05/10/2018, and on receipt of representations from vendors, following amendment/changes in the specifications are hereby notified. These are:

S.no.	Specifications in existing Tender	Amended specifications
1	<p>➤ Equipment</p> <p>➤ BP Apparatus (aneroid dial type ISI or BIS mark)</p> <p>➤ ECG Machine 12 lead with a warranty of 02 years and CMC 03 years</p> <p>➤ Mini Lab (accuster etc.)</p> <p>➤ Refrigerator 150- 200 lt minimum 3 star BEE rating with standard warranty reputed make (whirlpool, LG, Samsung etc) with bottle dispenser .</p>	<p>➤ Equipment</p> <p>➤ No Change</p> <p>➤ No Change</p> <p>➤ Mini Lab : a). Semi-automatic Chemistry analyzer; b) Incubator; c) Centrifuge (at least 12 tubes); d) Sahli's Hemoglobinometer; e) Glucometer; etc.</p> <p>➤ Refrigerator 50- 200 lt minimum 3 star BEE rating with standard warranty reputed make (whirlpool, LG, Samsung etc) with bottle dispenser (Bottle Dispenser is Optional);</p>
2.	<p>Ophthalmology Examination Unit with list of instruments attached</p> <p>➤ Snellen's Acuity Chart-Illuminated mirror-rotor test type drum with remote control, English, Hindi, Punjabi , illiterate C/E, Duochrome worth's Four Dot Point Etc.(Aurolab, Appaswamy , Medele)etc.</p>	<p>Ophthalmology Examination Unit with list of instruments attached</p> <p>Snellen's Acuity Chart: LED Digital Visual acuity Chart (Which can be used even at distances of 1 /2/3 meters, while illuminated Mirror -rotor type requires at least a distance of 6 meter.) Other details remain same. All Lab. Equipments With Standard warranty unless otherwise mentioned in T&C.</p>

The other details of Specification remain unchanged.

This is for information of all.


Addl. DHA (M&TB)

Copy to;

1. Director IT to kindly upload in the website of SDMC & also be shown as 'flash';
2. O/c

Copy for information:

1. PS to Commissioner for information of Commissioner;
2. Addl. Commissioner(Health) & CA-cum-FA
3. DC(health)
4. DCA(F&G)